

LENTEK™ Deionized Water

Ultra-purified water with dissolved minerals and impurities removed. Helps prevent scale buildup, spotting, and contamination—ensuring precision, consistency, and peak performance.



FINAL TESTING REQUIREMENTS

Test and Test Condition	Limit	Unit	Method	Note
Total Hardness	10 Max	ppm	ASTM D511	1
Conductivity	25 Max	mmho	ASTM D1125	
Chloride	10 Max	ppm	ASTM D512	
pH	5.0 – 9.0		ASTM E70	

Final Testing Requirements Notes:

- 1 Total Hardness can also be measured by ASTM D1126-86 (92). Same limits.

GENERAL NOTES

- Conductivity measurement unit as mmho is defined as micromho's per cm. The test method states that this is equivalent to s/cm which is siemens per centimeter.

PACKAGING

PART N°	SIZE	CASE / PALLET
9990502	5 Gal	1 / 24
9990501	55 Gal	1 / 4
9990500	265 Gal	1 / 1

SPECIFICATION NUMBER

- 000000053542

Read precautionary information and material safety sheets. This product is shipped in compliance with applicable laws and regulations regarding classification, packaging, shipping and labeling.